

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062518\
 Data File : VX002888.D
 Acq On : 26 Jun 2018 04:19
 Operator : JC/MD
 Sample : J3693-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-25

Quant Time: Jun 27 07:47:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254937	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	352475	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	331469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	190167	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	172425	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
35) Dibromofluoromethane	5.51	113	136857	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	8.72	98	504269	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
62) 4-Bromofluorobenzene	11.14	95	179949	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	1540	6.70	ug/l #	84
16) Acetone	2.45	43	1653	1.13	ug/l	95
19) Methyl tert-butyl Ether	3.20	73	6197	0.69	ug/l	92
65) Chlorobenzene	10.14	112	2000	0.24	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	1736	0.25	ug/l #	38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062518\
Data File : VX002888.D
Acq On : 26 Jun 2018 04:19
Operator : JC/MD
Sample : J3693-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-25

Quant Time: Jun 27 07:47:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

