

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062518\
 Data File : VX002862.D
 Acq On : 25 Jun 2018 13:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 25 14:56:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243572	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	337142	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174017	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161759	48.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.10%	
35) Dibromofluoromethane	5.51	113	126929	47.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.08%	
50) Toluene-d8	8.72	98	470880	48.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.60%	
62) 4-Bromofluorobenzene	11.14	95	167608	44.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.16%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062518\
Data File : VX002862.D
Acq On : 25 Jun 2018 13:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 25 14:56:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. 0.00 min

Lab File: VX002862.D

Acq: 25 Jun 2018 13:33

Instrument :

MSVOA_X

ClientSampleId :


IBLK

Tot Ion:168 Resp: 243572

Ion Ratio Lower Upper

168 100

99 49.0 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000

60000



40000

20000

0

Time-->

5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 48.05 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX002862.D


Acq: 25 Jun 2018 13:33

Tgt Ion: 65 Resp: 161759

Ion Ratio Lower Upper

65 100

67 50.3 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000


40000


20000

0

Time-->

6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX002862.D

Acq: 25 Jun 2018 13:33

Instrument :

MSVOA_X

ClientSampleId :

IBLK


Tot Ion:114 Resp: 337142

Ion Ratio Lower Upper

114 100

63 20.3 0.0 41.4




88 15.2 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.54 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX002862.D

Acq: 25 Jun 2018 13:33


Tgt Ion:113 Resp: 126929

Ion Ratio Lower Upper

113 100

111 101.1 81.4 122.0



192 24.3 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.30 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002862.D

Acq: 25 Jun 2018 13:33

Instrument :

MSVOA_X

ClientSampleId :


IBLK

Tot Ion: 98 Resp: 470880

Ion Ratio Lower Upper




98 100

100 63.8 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 44.58 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX002862.D

Acq: 25 Jun 2018 13:33


Tgt Ion: 95 Resp: 167608

Ion Ratio Lower Upper

95 100

174 97.1 0.0 187.4



176 93.1 0.0 181.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX002862.D
Acq: 25 Jun 2018 13:33

Instrument :
MSVOA_X
ClientSampled :
IBLK

Tot Ion:117 Resp: 306914
Ion Ratio Lower Upper
117 100
82 53.1 45.0 67.4
119 33.0 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.01 min
Lab File: VX002862.D
Acq: 25 Jun 2018 13:33

Tgt Ion:152 Resp: 174017
Ion Ratio Lower Upper
152 100
115 55.4 40.1 120.2
150 155.4 0.0 347.2

