

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062519\
 Data File : VX010440.D
 Acq On : 25 Jun 2019 17:52
 Operator : JC/SP
 Sample : K3519-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-19-LF2MW4-9001

Quant Time: Jun 26 02:01:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	421686	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	384761	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176227	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	129396	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
35) Dibromofluoromethane	5.50	113	129386	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
50) Toluene-d8	8.71	98	498174	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	11.13	95	166606	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062519\
Data File : VX010440.D
Acq On : 25 Jun 2019 17:52
Operator : JC/SP
Sample : K3519-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-19-LF2MW4-9001

Quant Time: Jun 26 02:01:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

