

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022915.D
 Acq On : 25 Jun 2021 12:05
 Operator : JC/MD
 Sample : VX0625WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 26 03:52:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	129266	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51981	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55779	46.56	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.12%
35) Dibromofluoromethane	5.391	113	39566	48.20	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.40%
50) Toluene-d8	8.653	98	157419	48.93	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.86%
62) 4-Bromofluorobenzene	11.085	95	58191	48.47	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.94%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.172	85	17869	19.19	ug/l 99
3) Chloromethane	1.294	50	17191	16.81	ug/l 99
4) Vinyl Chloride	1.374	62	16697	17.69	ug/l 99
5) Bromomethane	1.605	94	9496	18.27	ug/l 98
6) Chloroethane	1.685	64	10405	17.84	ug/l 94
7) Trichlorofluoromethane	1.892	101	25111	18.32	ug/l 100
8) Diethyl Ether	2.142	74	9214	17.41	ug/l 83
9) 1,1,2-Trichlorotrifluo...	2.331	101	17002	19.81	ug/l 88
10) Methyl Iodide	2.453	142	12869	15.38	ug/l 93
11) Tert butyl alcohol	2.971	59	19719	78.98	ug/l 97
12) 1,1-Dichloroethene	2.319	96	14984	17.50	ug/l 89
13) Acrolein	2.239	56	13835	89.10	ug/l 99
14) Allyl chloride	2.666	41	27515	16.91	ug/l 90
15) Acrylonitrile	3.075	53	52989	92.84	ug/l 100
16) Acetone	2.386	43	46924	88.36	ug/l 93
17) Carbon Disulfide	2.514	76	32072	16.29	ug/l 98
18) Methyl Acetate	2.709	43	24127	18.46	ug/l # 88
19) Methyl tert-butyl Ether	3.123	73	55542	18.09	ug/l # 89
20) Methylene Chloride	2.794	84	18432	17.74	ug/l # 84
21) trans-1,2-Dichloroethene	3.093	96	16671	17.55	ug/l 85
22) Diisopropyl ether	3.769	45	62324	18.42	ug/l 97
23) Vinyl Acetate	3.733	43	259305	91.34	ug/l 95
24) 1,1-Dichloroethane	3.617	63	32930	18.21	ug/l 99
25) 2-Butanone	4.568	43	74730	90.49	ug/l 90
26) 2,2-Dichloropropane	4.483	77	24256	17.68	ug/l 97
27) cis-1,2-Dichloroethene	4.495	96	19584	18.57	ug/l 96
28) Bromochloromethane	4.903	49	15534	18.24	ug/l # 77
29) Tetrahydrofuran	5.019	42	48647	91.58	ug/l # 87
30) Chloroform	5.099	83	33284	18.30	ug/l 97
31) Cyclohexane	5.477	56	29962	17.72	ug/l # 92
32) 1,1,1-Trichloroethane	5.391	97	25804	17.66	ug/l 95
36) 1,1-Dichloropropene	5.696	75	24206	17.63	ug/l 98
37) Ethyl Acetate	4.727	43	29660	18.65	ug/l # 91
38) Carbon Tetrachloride	5.690	117	20918	18.16	ug/l 93
39) Methylcyclohexane	7.385	83	29676	18.53	ug/l 95
40) Benzene	6.043	78	73297	18.51	ug/l 100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022915.D
 Acq On : 25 Jun 2021 12:05
 Operator : JC/MD
 Sample : VX0625WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 26 03:52:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	15223	17.37	ug/l	91
42) 1,2-Dichloroethane	6.092	62	27860	19.13	ug/l	96
43) Isopropyl Acetate	6.348	43	46586	18.22	ug/l #	92
44) Trichloroethene	7.129	130	17431	18.57	ug/l	89
45) 1,2-Dichloropropane	7.440	63	19029	18.63	ug/l	99
46) Dibromomethane	7.586	93	12352	18.23	ug/l #	85
47) Bromodichloromethane	7.824	83	21103	17.36	ug/l	96
48) Methyl methacrylate	7.696	41	21294	17.99	ug/l	87
49) 1,4-Dioxane	7.665	88	8670	340.34	ug/l	92
51) 4-Methyl-2-Pentanone	8.580	43	149407	93.88	ug/l	89
52) Toluene	8.720	92	44270	18.50	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	22807	16.64	ug/l	100
54) cis-1,3-Dichloropropene	8.372	75	26773	17.20	ug/l #	91
55) 1,1,2-Trichloroethane	9.159	97	18211	18.96	ug/l	97
56) Ethyl methacrylate	9.122	69	28138	17.63	ug/l #	83
57) 1,3-Dichloropropane	9.311	76	31910	18.28	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	75745	89.39	ug/l	95
59) 2-Hexanone	9.433	43	113297	93.84	ug/l	86
60) Dibromochloromethane	9.525	129	13071	16.53	ug/l	94
61) 1,2-Dibromoethane	9.616	107	17721	18.17	ug/l	100
64) Tetrachloroethene	9.275	164	15836	18.90	ug/l #	87
65) Chlorobenzene	10.085	112	45997	18.32	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	14554	17.93	ug/l	98
67) Ethyl Benzene	10.195	91	84785	18.34	ug/l	99
68) m/p-Xylenes	10.305	106	64028	37.29	ug/l	96
69) o-Xylene	10.646	106	30606	18.31	ug/l	95
70) Styrene	10.659	104	51420	18.54	ug/l	97
71) Bromoform	10.805	173	7174	15.80	ug/l #	96
73) Isopropylbenzene	10.963	105	81676	18.57	ug/l	98
74) N-amyl acetate	10.848	43	38411	16.89	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	27743	18.36	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	29418	21.43	ug/l	92
77) Bromobenzene	11.201	156	16632	18.18	ug/l	81
78) n-propylbenzene	11.305	91	100328	18.85	ug/l	95
79) 2-Chlorotoluene	11.366	91	58457	18.19	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	68431	18.95	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.024	75	6223	17.16	ug/l #	73
82) 4-Chlorotoluene	11.457	91	68386	18.42	ug/l	95
83) tert-Butylbenzene	11.719	119	61748	18.10	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	69563	19.14	ug/l	98
85) sec-Butylbenzene	11.890	105	87805	19.09	ug/l	97
86) p-Isopropyltoluene	12.012	119	70444	19.27	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	33277	18.86	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	33439	18.12	ug/l	96
89) n-Butylbenzene	12.335	91	66631	19.17	ug/l	95
90) Hexachloroethane	12.542	117	9278	16.42	ug/l	74
91) 1,2-Dichlorobenzene	12.335	146	32690	18.98	ug/l	94
92) 1,2-Dibromo-3-Chloropr...	12.945	75	4759	16.25	ug/l	76
93) 1,2,4-Trichlorobenzene	13.591	180	18801	19.68	ug/l	94
94) Hexachlorobutadiene	13.725	225	7533	20.13	ug/l	94
95) Naphthalene	13.780	128	69912	17.79	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	18754	19.49	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022915.D
Acq On : 25 Jun 2021 12:05
Operator : JC/MD
Sample : VX0625WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Jun 26 03:52:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 26 03:52:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:168 Resp: 66010
Ion Ratio Lower Upper
168 100
99 64.8 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 19.19 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 85 Resp: 17869
Ion Ratio Lower Upper
85 100
87 33.2 16.4 49.2









#3
Chloromethane
Concen: 16.81 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 17191
Ion Ratio Lower Upper
50 100
52 33.4 26.4 39.6



#4
Vinyl Chloride
Concen: 17.69 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 62 Resp: 16697
Ion Ratio Lower Upper
62 100
64 31.3 25.7 38.5









#5
Bromomethane
Concen: 18.27 ug/l
RT: 1.605 min Scan# 85
Delta R.T. 0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 9496
Ion Ratio Lower Upper
94 100
96 93.5 76.4 114.6



#6
Chloroethane
Concen: 17.84 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.007 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 64 Resp: 10405
Ion Ratio Lower Upper
64 100
66 28.2 25.4 38.0









#7
Trichlorofluoromethane
Concen: 18.32 ug/l
RT: 1.892 min Scan# 132
Delta R.T. 0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 25111
Ion Ratio Lower Upper
101 100
103 66.4 53.2 79.8



#8
Diethyl Ether
Concen: 17.41 ug/l
RT: 2.142 min Scan# 173
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 74 Resp: 9214
Ion Ratio Lower Upper
74 100
45 104.9 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.81 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 17002
Ion Ratio Lower Upper
101 100
85 46.9 34.9 52.3
151 63.5 61.4 92.2



#10
Methyl Iodide
Concen: 15.38 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:142 Resp: 12869
Ion Ratio Lower Upper
142 100
127 47.4 33.3 49.9
141 13.5 11.4 17.0









#11

Tert butyl alcohol
Concen: 78.98 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 59 Resp: 19719
Ion Ratio Lower Upper
59 100
57 11.1 7.8 11.8




#12

1,1-Dichloroethene
Concen: 17.50 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.001 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 96 Resp: 14984
Ion Ratio Lower Upper
96 100
61 175.9 124.6 187.0
98 64.5 50.6 76.0









#13
Acrolein
Concen: 89.10 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 13835
Ion Ratio Lower Upper
56 100
55 71.4 56.4 84.6



#14
Allyl chloride
Concen: 16.91 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 41 Resp: 27515
Ion Ratio Lower Upper
41 100
39 72.7 51.6 77.4
76 31.9 30.5 45.7









#15
Acrylonitrile
Concen: 92.84 ug/l
RT: 3.075 min Scan# 326
Delta R.T. 0.001 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 52989
Ion Ratio Lower Upper
53 100
52 81.8 65.3 97.9
51 35.6 28.2 42.2



#16
Acetone
Concen: 88.36 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 43 Resp: 46924
Ion Ratio Lower Upper
43 100
58 30.6 27.8 41.8









#17
Carbon Disulfide
Concen: 16.29 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 32072
Ion Ratio Lower Upper
76 100
78 9.7 7.2 10.8



#18
Methyl Acetate
Concen: 18.46 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 43 Resp: 24127
Ion Ratio Lower Upper
43 100
74 21.1 21.9 32.9#









#19
Methyl tert-butyl Ether
Concen: 18.09 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 55542
Ion Ratio Lower Upper
73 100
57 26.8 17.1 25.7#



#20
Methylene Chloride
Concen: 17.74 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 84 Resp: 18432
Ion Ratio Lower Upper
84 100
49 139.2 88.6 132.8#
51 40.3 27.0 40.6
86 62.9 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 17.55 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 96 Resp: 16671
Ion Ratio Lower Upper
96 100
61 158.0 107.7 161.5
98 68.0 50.4 75.6



#22
Diisopropyl ether
Concen: 18.42 ug/l
RT: 3.769 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 45 Resp: 62324
Ion Ratio Lower Upper
45 100
43 72.5 58.0 87.0
87 25.5 25.4 38.0
59 11.7 10.2 15.2









#23
 Vinyl Acetate
 Concen: 91.34 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. -0.000 min
 Lab File: VX022915.D
 Acq: 25 Jun 2021 12:05




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 259305
 Ion Ratio Lower Upper
 43 100
 86 10.6 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 18.21 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX022915.D
 Acq: 25 Jun 2021 12:05

Tgt Ion: 63 Resp: 32930
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.8 11.2
 100 4.0 2.5 7.5





#25

2-Butanone

Concen: 90.49 ug/l

RT: 4.568 min Scan# 571

Delta R.T. 0.000 min

Lab File: VX022915.D





Acq: 25 Jun 2021 12:05

Tgt Ion: 43 Resp: 74730

Ion Ratio Lower Upper

43 100

72 24.1 23.7 35.5



#26

2,2-Dichloropropane

Concen: 17.68 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX022915.D




Acq: 25 Jun 2021 12:05


Tgt Ion: 77 Resp: 24256

Ion Ratio Lower Upper

77 100

97 21.7 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 18.57 ug/l

RT: 4.495 min Scan# 559





Delta R.T. -0.000 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion: 96 Resp: 19584

Ion	Ratio	Lower	Upper
96	100		
61	150.4	0.0	285.6
98	64.4	0.0	129.6



#28

Bromochloromethane

Concen: 18.24 ug/l

RT: 4.903 min Scan# 626




Delta R.T. 0.000 min


Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion: 49 Resp: 15534

Ion	Ratio	Lower	Upper
49	100		
129	0.9	0.0	4.0
130	64.0	68.2	102.2#





#29

Tetrahydrofuran

Concen: 91.58 ug/l

RT: 5.019 min Scan# 645





Delta R.T. 0.000 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion:	42	Resp:	48647
Ion	Ratio	Lower	Upper
42	100		
72	43.4	41.7	62.5
71	38.3	38.4	57.6#

Instrument :
MSVOA_X
ClientSampled :



#30

Chloroform

Concen: 18.30 ug/l




RT: 5.099 min Scan# 658


Delta R.T. -0.006 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05


Tgt Ion:	83	Resp:	33284
Ion	Ratio	Lower	Upper
83	100		
85	62.3	52.0	78.0








#31
Cyclohexane
Concen: 17.72 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.001 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 29962
Ion Ratio Lower Upper
56 100
69 25.3 25.3 37.9#
84 83.1 71.4 107.0




#32
1,1,1-Trichloroethane
Concen: 17.66 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05



Tgt Ion: 97 Resp: 25804
Ion Ratio Lower Upper
97 100
99 65.5 51.5 77.3
61 49.6 34.2 51.4









#33
1,2-Dichloroethane-d4
Concen: 46.56 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 65 Resp: 55779
Ion Ratio Lower Upper
65 100
67 53.0 0.0 105.4



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 114 Resp: 129266
Ion Ratio Lower Upper
114 100
63 22.3 0.0 40.8
88 15.6 0.0 32.8





#35

Dibromofluoromethane

Concen: 48.20 ug/l

RT: 5.391 min Scan# 706





Delta R.T. -0.006 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion:	113	Resp:	39566
Ion	Ratio	Lower	Upper
113	100		
111	99.6	81.9	122.9
192	15.3	16.7	25.1#

Instrument :
MSVOA_X
ClientSampleId :



#36

1,1-Dichloropropene

Concen: 17.63 ug/l




RT: 5.696 min Scan# 756


Delta R.T. -0.006 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion:	75	Resp:	24206
Ion	Ratio	Lower	Upper
75	100		
110	34.1	18.1	54.4
77	31.6	25.3	37.9









#37
Ethyl Acetate
Concen: 18.65 ug/l
RT: 4.727 min Scan# 597
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	43	Resp:	29660
Ion Ratio	Lower	Upper	
43	100		
61	12.3	12.8	19.2#
70	9.7	10.5	15.7#



#38
Carbon Tetrachloride
Concen: 18.16 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:	117	Resp:	20918
Ion Ratio	Lower	Upper	
117	100		
119	101.6	76.2	114.2
121	25.7	24.8	37.2





#39

Methylcyclohexane

Concen: 18.53 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05





Tgt Ion: 83 Resp: 29676

Ion Ratio Lower Upper

83 100

55 77.2 58.9 88.3

98 52.1 38.1 57.1



#40

Benzene

Concen: 18.51 ug/l

RT: 6.043 min Scan# 813

Delta R.T. -0.007 min

Lab File: VX022915.D




Acq: 25 Jun 2021 12:05


Tgt Ion: 78 Resp: 73297

Ion Ratio Lower Upper

78 100

77 23.4 18.9 28.3









#41
Methacrylonitrile
Concen: 17.37 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	41	Resp:	15223
Ion	Ratio	Lower	Upper
41	100		
39	58.5	42.5	63.7
67	69.7	65.0	97.6
52	30.8	25.3	37.9



#42
1,2-Dichloroethane
Concen: 19.13 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:	62	Resp:	27860
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	20.2









#43
Isopropyl Acetate
Concen: 18.22 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 46586
Ion Ratio Lower Upper
43 100
61 18.9 18.2 27.4
87 11.9 12.5 18.7#



#44
Trichloroethene
Concen: 18.57 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:130 Resp: 17431
Ion Ratio Lower Upper
130 100
95 109.7 0.0 198.0





#45

1,2-Dichloropropane

Concen: 18.63 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.001 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion: 63 Resp: 19029

Ion Ratio Lower Upper





63 100

65 31.5 25.7 38.5

Instrument :

MSVOA_X

ClientSampled :



#46

Dibromomethane

Concen: 18.23 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05




Tgt Ion: 93 Resp: 12352


Ion Ratio Lower Upper

93 100

95 83.1 66.8 100.2

174 76.7 83.1 124.7#









#47
Bromodichloromethane
Concen: 17.36 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 21103
Ion Ratio Lower Upper
83 100
85 60.6 51.4 77.0
127 7.5 6.8 10.2



#48
Methyl methacrylate
Concen: 17.99 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 41 Resp: 21294
Ion Ratio Lower Upper
41 100
69 81.5 76.0 114.0
39 65.0 44.6 67.0









#49
1,4-Dioxane
Concen: 340.34 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 88 Resp: 8670
Ion Ratio Lower Upper
88 100
43 33.7 23.2 34.8
58 74.6 55.0 82.4



#50
Toluene-d8
Concen: 48.93 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 98 Resp: 157419
Ion Ratio Lower Upper
98 100
100 64.6 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 93.88 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 149407
Ion Ratio Lower Upper
43 100
58 36.8 35.4 53.0



#52
Toluene
Concen: 18.50 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 92 Resp: 44270
Ion Ratio Lower Upper
92 100
91 173.3 137.1 205.7









#53
 t-1,3-Dichloropropene
 Concen: 16.64 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. -0.000 min
 Lab File: VX022915.D
 Acq: 25 Jun 2021 12:05




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 75 Resp: 22807
 Ion Ratio Lower Upper
 75 100
 77 31.0 24.9 37.3



#54
 cis-1,3-Dichloropropene
 Concen: 17.20 ug/l
 RT: 8.372 min Scan# 1195
 Delta R.T. 0.000 min
 Lab File: VX022915.D
 Acq: 25 Jun 2021 12:05

Tgt Ion: 75 Resp: 26773
 Ion Ratio Lower Upper
 75 100
 77 30.9 25.7 38.5
 39 53.0 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 18.96 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 97 Resp: 18211




Ion	Ratio	Lower	Upper
97	100		
83	84.6	69.3	103.9
85	58.3	44.3	66.5
99	60.1	50.3	75.5




#56
Ethyl methacrylate
Concen: 17.63 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 69 Resp: 28138

Ion	Ratio	Lower	Upper
69	100		
41	69.3	46.8	70.2
39	43.6	24.7	37.1





#57

1,3-Dichloropropane

Concen: 18.28 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VX022915.D





Acq: 25 Jun 2021 12:05

Tgt Ion: 76 Resp: 31910

Ion Ratio Lower Upper

76 100

78 31.9 26.2 39.2



#58

2-Chloroethyl Vinyl ether

Concen: 89.39 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX022915.D




Acq: 25 Jun 2021 12:05


Tgt Ion: 63 Resp: 75745

Ion Ratio Lower Upper

63 100

106 26.6 23.5 35.3









#59
2-Hexanone
Concen: 93.84 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 113297
Ion Ratio Lower Upper
43 100
58 51.1 30.9 92.8



#60
Dibromochloromethane
Concen: 16.53 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.001 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 129 Resp: 13071
Ion Ratio Lower Upper
129 100
127 81.6 38.4 115.2









#61
1,2-Dibromoethane
Concen: 18.17 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 17721
Ion Ratio Lower Upper
107 100
109 94.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.47 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 95 Resp: 58191
Ion Ratio Lower Upper
95 100
174 61.1 0.0 154.6
176 59.0 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 119242
Ion Ratio Lower Upper
117 100
82 62.1 47.4 71.2
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 18.90 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:164 Resp: 15836
Ion Ratio Lower Upper
164 100
166 125.8 103.4 155.2
129 113.6 70.8 106.2#
131 104.4 72.5 108.7









#65
Chlorobenzene
Concen: 18.32 ug/l
RT: 10.085 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 45997
Ion Ratio Lower Upper
112 100
114 31.0 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 17.93 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:131 Resp: 14554
Ion Ratio Lower Upper
131 100
133 93.0 47.7 143.1
119 66.2 32.1 96.5









#67
Ethyl Benzene
Concen: 18.34 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 84785
Ion Ratio Lower Upper
91 100
106 30.3 24.6 37.0



#68
m/p-Xylenes
Concen: 37.29 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 106 Resp: 64028
Ion Ratio Lower Upper
106 100
91 210.2 162.8 244.2









#69
o-Xylene
Concen: 18.31 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 30606
Ion Ratio Lower Upper
106 100
91 223.4 108.1 324.2



#70
Styrene
Concen: 18.54 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.001 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:104 Resp: 51420
Ion Ratio Lower Upper
104 100
78 52.9 40.0 60.0
103 53.5 43.4 65.0





#71

Bromoform

Concen: 15.80 ug/l

RT: 10.805 min Scan# 1594





Delta R.T. -0.000 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion:	173	Resp:	7174
Ion Ratio	Lower	Upper	
173	100		
175	50.3	23.9	71.8
254	0.0	0.1	0.1#

Instrument :
MSVOA_X
ClientSampleId :



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX022915.D

Acq: 25 Jun 2021 12:05

Tgt Ion:	152	Resp:	51981
Ion Ratio	Lower	Upper	
152	100		
115	77.3	41.1	123.3
150	165.9	0.0	349.0









#73
Isopropylbenzene
Concen: 18.57 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 81676
Ion Ratio Lower Upper
105 100
120 25.5 13.2 39.6



#74
N-ethyl acetate
Concen: 16.89 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.001 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 43 Resp: 38411
Ion Ratio Lower Upper
43 100
70 39.5 39.0 58.6
55 23.9 21.5 32.3
61 22.7 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 18.36 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 27743
Ion Ratio Lower Upper
83 100
131 8.1 5.1 15.2
85 63.5 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 21.43 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 75 Resp: 29418
Ion Ratio Lower Upper
75 100
77 33.4 19.1 57.5









#77
Bromobenzene
Concen: 18.18 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 16632
Ion Ratio Lower Upper
156 100
77 204.6 82.8 248.6
158 96.7 49.1 147.4



#78
n-propylbenzene
Concen: 18.85 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 91 Resp: 100328
Ion Ratio Lower Upper
91 100
120 20.5 11.5 34.5









#79
2-Chlorotoluene
Concen: 18.19 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 58457
Ion Ratio Lower Upper
91 100
126 31.5 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 18.95 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 105 Resp: 68431
Ion Ratio Lower Upper
105 100
120 46.8 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 17.16 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 6223
Ion Ratio Lower Upper
75 100
53 126.8 74.5 111.7#
89 39.9 37.9 56.9



#82
4-Chlorotoluene
Concen: 18.42 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 91 Resp: 68386
Ion Ratio Lower Upper
91 100
126 27.0 14.9 44.5









#83
tert-Butylbenzene
Concen: 18.10 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:119 Resp: 61748
Ion Ratio Lower Upper
119 100
91 65.5 27.9 83.7
134 23.2 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 19.14 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:105 Resp: 69563
Ion Ratio Lower Upper
105 100
120 43.5 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 19.09 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.006 min
 Lab File: VX022915.D
 Acq: 25 Jun 2021 12:05




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 87805
 Ion Ratio Lower Upper
 105 100
 134 18.4 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 19.27 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX022915.D
 Acq: 25 Jun 2021 12:05

Tgt Ion:119 Resp: 70444
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.7
 91 27.0 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 18.86 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 33277
Ion Ratio Lower Upper
146 100
111 47.0 20.2 60.5
148 64.1 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 18.12 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:146 Resp: 33439
Ion Ratio Lower Upper
146 100
111 45.2 19.6 58.8
148 63.3 31.6 95.0









#89
n-Butylbenzene
Concen: 19.17 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 66631
Ion Ratio Lower Upper
91 100
92 51.5 27.4 82.0
134 23.5 13.2 39.5



#90
Hexachloroethane
Concen: 16.42 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 117 Resp: 9278
Ion Ratio Lower Upper
117 100
201 57.6 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 18.98 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 32690
Ion Ratio Lower Upper
146 100
111 48.6 20.8 62.2
148 61.8 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.25 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion: 75 Resp: 4759
Ion Ratio Lower Upper
75 100
155 69.5 45.3 135.9
157 87.5 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 19.68 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:180 Resp: 18801
Ion Ratio Lower Upper
180 100
182 90.1 47.7 143.1
145 35.2 15.0 45.0



#94
Hexachlorobutadiene
Concen: 20.13 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:225 Resp: 7533
Ion Ratio Lower Upper
225 100
223 66.2 31.1 93.2
227 57.5 31.6 95.0









#95
Naphthalene
Concen: 17.79 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:128 Resp: 69912
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 10.8 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.49 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022915.D
Acq: 25 Jun 2021 12:05

Tgt Ion:180 Resp: 18754
Ion Ratio Lower Upper
180 100
182 90.8 47.6 142.9
145 35.7 16.1 48.3

