

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022913.D
 Acq On : 25 Jun 2021 11:15
 Operator : JC/MD
 Sample : VX0625MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0625MBL01

Quant Time: Jun 26 03:51:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	103609	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	92065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34318	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45300	47.255	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.520%
35) Dibromofluoromethane	5.391	113	29901	45.446	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.900%
50) Toluene-d8	8.653	98	129630	50.272	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.540%
62) 4-Bromofluorobenzene	11.085	95	43356	45.052	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022913.D
Acq On : 25 Jun 2021 11:15
Operator : JC/MD
Sample : VX0625MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0625MBL01

Quant Time: Jun 26 03:51:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

