

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022923.D
 Acq On : 25 Jun 2021 15:16
 Operator : JC/MD
 Sample : M2803-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW201D-20210619DL

Quant Time: Jun 26 03:53:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	51267	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.769	114	107053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	48303	51.917	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.840%	
35) Dibromofluoromethane	5.397	113	34021	50.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.080%	
50) Toluene-d8	8.653	98	134862	50.618	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.240%	
62) 4-Bromofluorobenzene	11.085	95	49254	49.534	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.060%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	1683	2.327	ug/l	98
7) Trichlorofluoromethane	1.892	101	41300	38.790	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.331	101	683	1.025	ug/l #	88
12) 1,1-Dichloroethene	2.325	96	8074	12.140	ug/l	89
30) Chloroform	5.099	83	1686	1.194	ug/l #	76
32) 1,1,1-Trichloroethane	5.397	97	1770	1.560	ug/l #	83
44) Trichloroethene	7.129	130	27700	35.624	ug/l	91
64) Tetrachloroethene	9.275	164	22621	32.108	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022923.D
Acq On : 25 Jun 2021 15:16
Operator : JC/MD
Sample : M2803-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW201D-20210619DL

Quant Time: Jun 26 03:53:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

