

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022924.D
 Acq On : 25 Jun 2021 15:39
 Operator : JC/MD
 Sample : M2803-14DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20210621DL

Quant Time: Jun 26 03:54:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	49991	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.769	114	101029	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	93963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	45678	50.349	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.700%
35) Dibromofluoromethane	5.397	113	31622	49.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.580%
50) Toluene-d8	8.653	98	127392	50.665	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.340%
62) 4-Bromofluorobenzene	11.085	95	44536	47.460	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.920%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	827	1.036	ug/l	93
44) Trichloroethene	7.135	130	37649	51.306	ug/l	91
64) Tetrachloroethene	9.274	164	1899	2.876	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022924.D
Acq On : 25 Jun 2021 15:39
Operator : JC/MD
Sample : M2803-14DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20210621DL

Quant Time: Jun 26 03:54:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

