

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022925.D
 Acq On : 25 Jun 2021 16:03
 Operator : JC/MD
 Sample : M2803-15DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20210621DL

Quant Time: Jun 26 03:54:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51508	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	107648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	100178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47112	50.400	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.800%	
35) Dibromofluoromethane	5.397	113	33716	49.322	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.640%	
50) Toluene-d8	8.653	98	134303	50.130	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.260%	
62) 4-Bromofluorobenzene	11.085	95	49523	49.529	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.060%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	23840	35.595	ug/l	89
44) Trichloroethene	7.129	130	15180	19.415	ug/l	92
64) Tetrachloroethene	9.275	164	1494	2.122	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022925.D
Acq On : 25 Jun 2021 16:03
Operator : JC/MD
Sample : M2803-15DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20210621DL

Quant Time: Jun 26 03:54:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

