

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022929.D
 Acq On : 25 Jun 2021 17:37
 Operator : JC/MD
 Sample : M2838-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20210621DL

Quant Time: Jun 26 03:54:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51509	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	106164	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	99649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41975	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47914	51.257	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.520%
35) Dibromofluoromethane	5.397	113	33821	50.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.340%
50) Toluene-d8	8.653	98	135125	51.141	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.280%
62) 4-Bromofluorobenzene	11.085	95	49286	49.981	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.960%
Target Compounds						
						Qvalue
44) Trichloroethene	7.135	130	41411	53.703	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022929.D
Acq On : 25 Jun 2021 17:37
Operator : JC/MD
Sample : M2838-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20210621DL

Quant Time: Jun 26 03:54:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

