

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022930.D
 Acq On : 25 Jun 2021 18:00
 Operator : JC/MD
 Sample : M2803-18DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20210621DL

Quant Time: Jun 26 03:55:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50449	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	105923	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47800	52.209	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.420%
35) Dibromofluoromethane	5.403	113	32537	48.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.740%
50) Toluene-d8	8.653	98	132276	50.177	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.360%
62) 4-Bromofluorobenzene	11.085	95	48722	49.522	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.040%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	823	1.255	ug/l	94
16) Acetone	2.386	43	565	1.392	ug/l	97
44) Trichloroethene	7.135	130	81862	106.403	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 26 03:55:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

