

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022931.D
 Acq On : 25 Jun 2021 18:24
 Operator : JC/MD
 Sample : M2838-11DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20210622DL

Quant Time: Jun 26 03:55:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	50756	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.769	114	103924	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	47239	51.285	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.560%
35) Dibromofluoromethane	5.397	113	32337	49.000	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	8.653	98	132407	51.193	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.380%
62) 4-Bromofluorobenzene	11.085	95	47150	48.846	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.700%
Target Compounds						
44) Trichloroethene	7.135	130	30462	40.356	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022931.D
Acq On : 25 Jun 2021 18:24
Operator : JC/MD
Sample : M2838-11DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20210622DL

Quant Time: Jun 26 03:55:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

