

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022934.D
 Acq On : 25 Jun 2021 19:34
 Operator : JC/MD
 Sample : M2838-04DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20210622DL

Quant Time: Jun 26 03:55:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50655	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	105824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51553	56.080	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.160%	
35) Dibromofluoromethane	5.397	113	35560	52.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.840%	
50) Toluene-d8	8.653	98	143196	54.370	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.740%	
62) 4-Bromofluorobenzene	11.085	95	51732	52.630	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.260%	
Target Compounds						
44) Trichloroethene	7.135	130	43433	56.506	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022934.D
Acq On : 25 Jun 2021 19:34
Operator : JC/MD
Sample : M2838-04DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20210622DL

Quant Time: Jun 26 03:55:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

