

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022941.D
 Acq On : 25 Jun 2021 23:29
 Operator : JC/MD
 Sample : VX0625WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0625WBL02

Quant Time: Jun 26 03:57:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52152	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	108861	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	99440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43647	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48635	51.387	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.780%
35) Dibromofluoromethane	5.403	113	33142	47.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	8.653	98	134478	49.636	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.280%
62) 4-Bromofluorobenzene	11.085	95	50505	49.949	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022941.D
Acq On : 25 Jun 2021 23:29
Operator : JC/MD
Sample : VX0625WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL02

Quant Time: Jun 26 03:57:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

