

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022944.D
 Acq On : 26 Jun 2021 00:40
 Operator : JC/MD
 Sample : PB137293TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293TB

Quant Time: Jun 26 03:58:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50484	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	105411	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42771	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	47876	52.256	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.520%	
35) Dibromofluoromethane	5.397	113	33007	49.309	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.620%	
50) Toluene-d8	8.652	98	136312	51.959	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	103.920%	
62) 4-Bromofluorobenzene	11.085	95	49249	50.301	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	100.600%	
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	4833	11.899	ug/l	# 86
20) Methylene Chloride	2.800	84	3979	5.275	ug/l	99
43) Isopropyl Acetate	6.360	43	10161	4.873	ug/l	95
95) Naphthalene	13.780	128	972	2.191	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022944.D
Acq On : 26 Jun 2021 00:40
Operator : JC/MD
Sample : PB137293TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293TB

Quant Time: Jun 26 03:58:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

