

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022946.D
 Acq On : 26 Jun 2021 01:27
 Operator : JC/MD
 Sample : PB137293ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#02

Quant Time: Jun 26 03:58:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	48228	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.775	114	100085	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	45623	52.126	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.260%		
35) Dibromofluoromethane	5.397	113	31088	48.914	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.820%		
50) Toluene-d8	8.653	98	130421	52.359	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 104.720%		
62) 4-Bromofluorobenzene	11.085	95	48639	52.321	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.640%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	6634	17.097	ug/l	# 87
20) Methylene Chloride	2.794	84	4911	6.706	ug/l	97
43) Isopropyl Acetate	6.348	43	10238	5.171	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022946.D
Acq On : 26 Jun 2021 01:27
Operator : JC/MD
Sample : PB137293ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#02

Quant Time: Jun 26 03:58:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

