

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022947.D
 Acq On : 26 Jun 2021 01:51
 Operator : JC/MD
 Sample : PB137293ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#03

Quant Time: Jun 26 03:58:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	49309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	104290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	103482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	46990	52.511	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.020%
35) Dibromofluoromethane	5.397	113	33054	49.910	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	136465	52.577	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.160%
62) 4-Bromofluorobenzene	11.085	95	51011	52.660	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.320%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6756	17.030	ug/l	89
20) Methylene Chloride	2.794	84	4989	6.666	ug/l #	85
43) Isopropyl Acetate	6.355	43	10311	4.998	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022947.D
Acq On : 26 Jun 2021 01:51
Operator : JC/MD
Sample : PB137293ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#03

Quant Time: Jun 26 03:58:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

