

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022950.D
 Acq On : 26 Jun 2021 03:01
 Operator : JC/MD
 Sample : PB137293ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#06

Quant Time: Jun 26 03:59:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51533	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	106549	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	105665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48076	51.406	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.820%	
35) Dibromofluoromethane	5.397	113	33284	49.192	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.380%	
50) Toluene-d8	8.653	98	139816	52.726	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.460%	
62) 4-Bromofluorobenzene	11.085	95	51262	51.797	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.600%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5867	14.151	ug/l	# 86
20) Methylene Chloride	2.800	84	4292	5.553	ug/l	# 81
43) Isopropyl Acetate	6.354	43	9965	4.728	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022950.D
Acq On : 26 Jun 2021 03:01
Operator : JC/MD
Sample : PB137293ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#06

Quant Time: Jun 26 03:59:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

