

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022951.D
 Acq On : 26 Jun 2021 03:25
 Operator : JC/MD
 Sample : PB137293ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#07

Quant Time: Jun 26 03:59:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52015	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	106976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	106956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48765	51.660	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.320%	
35) Dibromofluoromethane	5.397	113	33907	49.913	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.820%	
50) Toluene-d8	8.652	98	141726	53.233	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.460%	
62) 4-Bromofluorobenzene	11.085	95	53339	53.681	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.360%	
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	5589	13.355	ug/l	91
20) Methylene Chloride	2.800	84	4361	5.587	ug/l #	93
43) Isopropyl Acetate	6.354	43	10182	4.812	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022951.D
Acq On : 26 Jun 2021 03:25
Operator : JC/MD
Sample : PB137293ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#07

Quant Time: Jun 26 03:59:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

