

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022954.D
 Acq On : 26 Jun 2021 04:36
 Operator : JC/MD
 Sample : PB137293ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#10

Quant Time: Jun 26 07:37:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	47762	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	105929	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	103316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	50591	58.367	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.740%
35) Dibromofluoromethane	5.391	113	33166	49.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.653	98	138841	52.664	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.320%
62) 4-Bromofluorobenzene	11.085	95	49152	49.956	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.920%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6564	17.082	ug/l	97
20) Methylene Chloride	2.794	84	4744	6.551	ug/l #	86
43) Isopropyl Acetate	6.354	43	10565	5.042	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022954.D
Acq On : 26 Jun 2021 04:36
Operator : JC/MD
Sample : PB137293ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#10

Quant Time: Jun 26 07:37:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

