

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022956.D
 Acq On : 26 Jun 2021 05:23
 Operator : JC/MD
 Sample : PB137293ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#12

Quant Time: Jun 26 07:37:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51363	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	108381	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48537	52.071	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.140%
35) Dibromofluoromethane	5.397	113	33480	48.645	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.300%
50) Toluene-d8	8.653	98	140690	52.158	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.320%
62) 4-Bromofluorobenzene	11.085	95	52306	51.959	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.920%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5648	13.668	ug/l	98
20) Methylene Chloride	2.794	84	4656	6.011	ug/l	95
43) Isopropyl Acetate	6.354	43	10384	4.844	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022956.D
Acq On : 26 Jun 2021 05:23
Operator : JC/MD
Sample : PB137293ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#12

Quant Time: Jun 26 07:37:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

