

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022958.D
 Acq On : 26 Jun 2021 06:10
 Operator : JC/MD
 Sample : PB137293ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137293ZHE#14

Quant Time: Jun 26 07:38:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51685	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	108989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	48635	51.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.700%
35) Dibromofluoromethane	5.397	113	34632	49.977	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.960%
50) Toluene-d8	8.653	98	143390	52.798	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.600%
62) 4-Bromofluorobenzene	11.085	95	53557	52.840	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.680%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	5690	13.683	ug/l	93
20) Methylene Chloride	2.794	84	5675	7.202	ug/l	92
43) Isopropyl Acetate	6.354	43	10637	4.928	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022958.D
Acq On : 26 Jun 2021 06:10
Operator : JC/MD
Sample : PB137293ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137293ZHE#14

Quant Time: Jun 26 07:38:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

