

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022912.D
 Acq On : 25 Jun 2021 10:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 26 03:51:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	52.115	-4.2	94	0.00
3 P	Chloromethane	50.000	45.971	8.1	90	0.00
4 C	Vinyl Chloride	50.000	49.360	1.3#	92	0.00
5 T	Bromomethane	50.000	46.693	6.6	95	0.00
6 T	Chloroethane	50.000	47.386	5.2	88	0.00
7 T	Trichlorofluoromethane	50.000	49.313	1.4	93	0.00
8 T	Diethyl Ether	50.000	47.722	4.6	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.615	-5.2	99	0.00
10 T	Methyl Iodide	50.000	41.497	17.0	82	0.00
11 T	Tert butyl alcohol	250.000	220.229	11.9	86	0.00
12 CM	1,1-Dichloroethene	50.000	49.543	0.9#	94	0.00
13 T	Acrolein	250.000	213.587	14.6	85	0.00
14 T	Allyl chloride	50.000	49.225	1.5	93	0.00
15 T	Acrylonitrile	250.000	250.678	-0.3	92	0.00
16 T	Acetone	250.000	298.167	-19.3	120	0.00
17 T	Carbon Disulfide	50.000	42.943	14.1	90	0.00
18 T	Methyl Acetate	50.000	47.730	4.5	93	0.00
19 T	Methyl tert-butyl Ether	50.000	50.351	-0.7	94	0.00
20 T	Methylene Chloride	50.000	48.522	3.0	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.094	1.8	95	0.00
22 T	Diisopropyl ether	50.000	51.113	-2.2	94	0.00
23 T	Vinyl Acetate	250.000	263.224	-5.3	94	0.00
24 P	1,1-Dichloroethane	50.000	50.505	-1.0	95	0.00
25 T	2-Butanone	250.000	271.384	-8.6	101	0.00
26 T	2,2-Dichloropropane	50.000	52.621	-5.2	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.156	-0.3	93	0.00
28 T	Bromochloromethane	50.000	50.182	-0.4	96	0.00
29 T	Tetrahydrofuran	250.000	247.323	1.1	90	0.00
30 C	Chloroform	50.000	50.089	-0.2#	95	0.00
31 T	Cyclohexane	50.000	50.260	-0.5	95	0.00
32 T	1,1,1-Trichloroethane	50.000	51.554	-3.1	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.434	5.1	96	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	51.337	-2.7	96	0.00
36 T	1,1-Dichloropropene	50.000	51.692	-3.4	96	0.00
37 T	Ethyl Acetate	50.000	51.421	-2.8	94	0.00
38 T	Carbon Tetrachloride	50.000	55.165	-10.3	95	-0.01
39 T	Methylcyclohexane	50.000	55.235	-10.5	99	0.00
40 TM	Benzene	50.000	52.603	-5.2	95	0.00
41 T	Methacrylonitrile	50.000	50.929	-1.9	95	0.00
42 TM	1,2-Dichloroethane	50.000	53.391	-6.8	94	0.00
43 T	Isopropyl Acetate	50.000	51.636	-3.3	92	0.00
44 TM	Trichloroethene	50.000	53.357	-6.7	97	0.00
45 C	1,2-Dichloropropane	50.000	52.484	-5.0#	94	0.00
46 T	Dibromomethane	50.000	52.238	-4.5	96	0.00
47 T	Bromodichloromethane	50.000	48.728	2.5	96	0.00
48 T	Methyl methacrylate	50.000	53.465	-6.9	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022912.D
 Acq On : 25 Jun 2021 10:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 26 03:51:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1041.537	-4.2	91	0.00
50 S	Toluene-d8	50.000	51.204	-2.4	96	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.480	-7.4	92	0.00
52 CM	Toluene	50.000	53.591	-7.2#	95	0.00
53 T	t-1,3-Dichloropropene	50.000	46.977	6.0	93	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.897	2.2	93	0.00
55 T	1,1,2-Trichloroethane	50.000	53.713	-7.4	95	0.00
56 T	Ethyl methacrylate	50.000	48.558	2.9	92	0.00
57 T	1,3-Dichloropropane	50.000	52.766	-5.5	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	254.388	-1.8	94	0.00
59 T	2-Hexanone	250.000	284.087	-13.6	98	0.00
60 T	Dibromochloromethane	50.000	47.854	4.3	94	0.00
61 T	1,2-Dibromoethane	50.000	52.494	-5.0	94	0.00
62 S	4-Bromofluorobenzene	50.000	51.491	-3.0	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	93	0.00
64 T	Tetrachloroethene	50.000	54.501	-9.0	98	0.00
65 PM	Chlorobenzene	50.000	52.547	-5.1	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.023	-8.0	92	0.00
67 C	Ethyl Benzene	50.000	54.187	-8.4#	96	0.00
68 T	m/p-Xylenes	100.000	106.477	-6.5	96	0.00
69 T	o-Xylene	50.000	53.201	-6.4	95	0.00
70 T	Styrene	50.000	54.060	-8.1	96	0.00
71 P	Bromoform	50.000	44.807	10.4	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	54.575	-9.2	96	0.00
74 T	N-ethyl acetate	50.000	48.061	3.9	91	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.301	-2.6	90	0.00
76 T	1,2,3-Trichloropropane	50.000	53.852	-7.7	95	0.00
77 T	Bromobenzene	50.000	53.364	-6.7	97	0.00
78 T	n-propylbenzene	50.000	56.007	-12.0	97	0.00
79 T	2-Chlorotoluene	50.000	53.493	-7.0	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	56.394	-12.8	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.999	4.0	96	0.00
82 T	4-Chlorotoluene	50.000	54.458	-8.9	99	0.00
83 T	tert-Butylbenzene	50.000	53.994	-8.0	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	56.227	-12.5	96	0.00
85 T	sec-Butylbenzene	50.000	56.228	-12.5	97	0.00
86 T	p-Isopropyltoluene	50.000	57.338	-14.7	99	0.00
87 T	1,3-Dichlorobenzene	50.000	54.806	-9.6	98	0.00
88 T	1,4-Dichlorobenzene	50.000	51.400	-2.8	96	0.00
89 T	n-Butylbenzene	50.000	58.958	-17.9	100	0.00
90 T	Hexachloroethane	50.000	48.706	2.6	98	0.00
91 T	1,2-Dichlorobenzene	50.000	53.557	-7.1	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.238	7.5	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	56.592	-13.2	100	0.00
94 T	Hexachlorobutadiene	50.000	59.176	-18.4	113	0.00
95 T	Naphthalene	50.000	51.058	-2.1	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022912.D
Acq On : 25 Jun 2021 10:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 26 03:51:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	56.937	-13.9	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6