

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
 Data File : VX022914.D
 Acq On : 25 Jun 2021 11:37
 Operator : JC/MD
 Sample : VX0625WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0625WBL01

Quant Time: Jun 26 03:52:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54762	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112060	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39066	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54302	54.640	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.280%
35) Dibromofluoromethane	5.391	113	35841	50.366	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.740%
50) Toluene-d8	8.653	98	149658	53.662	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.320%
62) 4-Bromofluorobenzene	11.085	95	51628	49.602	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062521\
Data File : VX022914.D
Acq On : 25 Jun 2021 11:37
Operator : JC/MD
Sample : VX0625WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL01

Quant Time: Jun 26 03:52:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37




Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL01


Tgt Ion:168 Resp: 54762
Ion Ratio Lower Upper
168 100
99 70.5 45.8 68.6#



#33
1,2-Dichloroethane-d4
Concen: 54.640 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37

Tgt Ion: 65 Resp: 54302
Ion Ratio Lower Upper
65 100
67 51.4 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37




Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL01


Tgt Ion:114 Resp: 112060
Ion Ratio Lower Upper
114 100
63 22.8 0.0 40.8
88 17.6 0.0 32.8



#35
Dibromofluoromethane
Concen: 50.366 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37

Tgt Ion:113 Resp: 35841
Ion Ratio Lower Upper
113 100
111 106.7 81.9 122.9
192 15.8 16.7 25.1#









#50
Toluene-d8
Concen: 53.662 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37




Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL01


Tgt Ion: 98 Resp: 149658
Ion Ratio Lower Upper
98 100
100 65.4 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 49.602 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37

Tgt Ion: 95 Resp: 51628
Ion Ratio Lower Upper
95 100
174 61.9 0.0 154.6
176 60.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37

Instrument :
MSVOA_X
ClientSampleId :
VX0625WBL01

Tgt Ion:117 Resp: 103206
Ion Ratio Lower Upper
117 100
82 63.2 47.4 71.2
119 32.0 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022914.D
Acq: 25 Jun 2021 11:37

Tgt Ion:152 Resp: 39066
Ion Ratio Lower Upper
152 100
115 68.5 41.1 123.3
150 159.6 0.0 349.0

