

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
 Data File : VX042025.D
 Acq On : 25 Jun 2024 12:26
 Operator : JC/MD
 Sample : P3052-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW8-20240624

Quant Time: Jun 26 05:35:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	149742	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232415	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76795	44.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.940%
35) Dibromofluoromethane	5.385	113	72609	44.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.060%
50) Toluene-d8	8.646	98	270260	46.650	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.300%
62) 4-Bromofluorobenzene	11.079	95	103497	48.463	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	1855	2.447	ug/l #	84
30) Chloroform	5.104	83	3814	1.510	ug/l	96
64) Tetrachloroethene	9.274	164	755	0.449	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
Data File : VX042025.D
Acq On : 25 Jun 2024 12:26
Operator : JC/MD
Sample : P3052-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW8-20240624

Quant Time: Jun 26 05:35:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

