

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
 Data File : VX042026.D
 Acq On : 25 Jun 2024 12:49
 Operator : JC/MD
 Sample : P3052-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20240624

Quant Time: Jun 26 05:36:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148108	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	227307	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74775	43.783	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.560%
35) Dibromofluoromethane	5.385	113	70792	43.892	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.780%
50) Toluene-d8	8.647	98	265080	46.784	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.560%
62) 4-Bromofluorobenzene	11.079	95	99164	47.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
Data File : VX042026.D
Acq On : 25 Jun 2024 12:49
Operator : JC/MD
Sample : P3052-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20240624

Quant Time: Jun 26 05:36:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

