

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
 Data File : VX042034.D
 Acq On : 25 Jun 2024 15:57
 Operator : JC/MD
 Sample : P3056-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-TB-20240620

Quant Time: Jun 26 05:40:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	136171	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	209699	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69581	44.313	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.620%
35) Dibromofluoromethane	5.391	113	66216	44.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.000%
50) Toluene-d8	8.646	98	245350	46.937	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.880%
62) 4-Bromofluorobenzene	11.079	95	92076	47.786	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
Data File : VX042034.D
Acq On : 25 Jun 2024 15:57
Operator : JC/MD
Sample : P3056-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-TB-20240620

Quant Time: Jun 26 05:40:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

