

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
 Data File : VX042037.D
 Acq On : 25 Jun 2024 17:07
 Operator : JC/MD
 Sample : P3056-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-GW-665-667

Quant Time: Jun 26 05:41:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133279	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	205856	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67826	44.132	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.260%
35) Dibromofluoromethane	5.385	113	64105	43.888	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.780%
50) Toluene-d8	8.647	98	241766	47.115	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.240%
62) 4-Bromofluorobenzene	11.079	95	89053	47.079	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.160%
Target Compounds						
16) Acetone	2.392	43	3484	5.163	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
Data File : VX042037.D
Acq On : 25 Jun 2024 17:07
Operator : JC/MD
Sample : P3056-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-GW-665-667

Quant Time: Jun 26 05:41:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

