

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
 Data File : VX042042.D
 Acq On : 25 Jun 2024 19:05
 Operator : JC/MD
 Sample : P3069-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A1285

Quant Time: Jun 26 05:44:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138679	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	210923	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107083	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68876	43.071	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.140%
35) Dibromofluoromethane	5.385	113	67182	44.890	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.780%
50) Toluene-d8	8.647	98	249242	47.405	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.820%
62) 4-Bromofluorobenzene	11.079	95	94459	48.738	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062524\
Data File : VX042042.D
Acq On : 25 Jun 2024 19:05
Operator : JC/MD
Sample : P3069-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A1285

Quant Time: Jun 26 05:44:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

