

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062618\
 Data File : VX002899.D
 Acq On : 26 Jun 2018 16:09
 Operator : JC/MD
 Sample : J3570-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 27 02:38:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	255345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	365086	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	337863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	190940	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	178549	50.59	ug/l	0.00
Spiked Amount			Recovery	=	101.18%	
35) Dibromofluoromethane	5.51	113	141114	48.81	ug/l	0.00
Spiked Amount			Recovery	=	97.62%	
50) Toluene-d8	8.72	98	513935	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	11.14	95	181017	44.46	ug/l	0.00
Spiked Amount			Recovery	=	88.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	116	Below Cal	#	43
3) Chloromethane	1.33	50	1095	Below Cal		91
5) Bromomethane	1.65	94	1943	Below Cal	#	71
6) Chloroethane	1.71	64	407	Below Cal	#	56
8) Diethyl Ether	2.19	74	407	0.23	ug/l	79
10) Methyl Iodide	2.52	142	3712	7.36	ug/l	# 94
11) Tert butyl alcohol	3.03	59	4804	6.83	ug/l	# 73
13) Acrolein	2.28	56	778	6.19	ug/l	90
15) Acrylonitrile	3.15	53	744	0.50	ug/l	# 92
16) Acetone	2.45	43	62096	42.48	ug/l	100
17) Carbon Disulfide	2.57	76	4047	0.60	ug/l	# 95
18) Methyl Acetate	2.78	43	1572	0.38	ug/l	# 88
20) Methylene Chloride	2.86	84	2016	0.71	ug/l	# 93
21) trans-1,2-Dichloroethene	3.17	96	574	0.21	ug/l	# 63
25) 2-Butanone	4.71	43	14150	6.52	ug/l	96
29) Tetrahydrofuran	5.17	42	715	0.51	ug/l	# 78
36) 1,1-Dichloropropene	5.67	75	17564	4.39	ug/l	# 50
38) Carbon Tetrachloride	5.67	117	23082	5.25	ug/l	# 16
49) 1,4-Dioxane	7.59	88	56	0.77	ug/l	# 63
51) 4-Methyl-2-Pentanone	8.65	43	1365	0.30	ug/l	100
59) 2-Hexanone	9.50	43	723	0.21	ug/l	100
76) 1,2,3-Trichloropropane	11.14	75	95765	27.69	ug/l	# 33
81) trans-1,4-Dichloro-2-buten	11.14	75	95765	78.30	ug/l	# 7
95) Naphthalene	13.84	128	3218	0.25	ug/l	# 91
96) 1,2,3-Trichlorobenzene	14.02	180	1033	0.21	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.00 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09


Instrument :
MSVOA_X
ClientSampled :

Tot Ion:168 Resp: 255345

Ion Ratio Lower Upper

168 100

99 51.7 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#2

Dichlorodifluoromethane

Concen: Below Cal

RT: 1.20 min Scan# 101

Delta R.T. 0.01 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 85 Resp: 116

Ion Ratio Lower Upper

85 100

87 0.0 16.0 47.9#



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0


150


100

50

0

Time--> 1.18 1.20 1.22





#3

Chloromethane

Concen: Below Cal

RT: 1.33 min Scan# 122

Delta R.T. 0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 50 Resp: 1095

Ion Ratio Lower Upper




50 100

52 37.2 25.7 38.5



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0



#5

Bromomethane

Concen: Below Cal

RT: 1.65 min Scan# 174

Delta R.T. 0.01 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 94 Resp: 1943

Ion Ratio Lower Upper



94 100


96 65.4 74.9 112.3#



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0





#6

Chloroethane

Concen: Below Cal

RT: 1.71 min Scan# 184

Delta R.T. -0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 64 Resp: 407

Ion Ratio Lower Upper

64 100

66 7.5 25.5 38.3#



#8

Diethyl Ether

Concen: 0.23 ug/l

RT: 2.19 min Scan# 263

Delta R.T. -0.00 min

Lab File: VX002899.D




Acq: 26 Jun 2018 16:09


Tgt Ion: 74 Resp: 407

Ion Ratio Lower Upper

74 100

45 85.0 53.1 159.4





#10

Methyl Iodide

Concen: 7.36 ug/l

RT: 2.52 min Scan# 317

Delta R.T. 0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:142 Resp: 3712

Ion Ratio Lower Upper

142 100

127 48.9 36.6 55.0




141 18.7 11.3 16.9#



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V



#11

Tert butyl alcohol

Concen: 6.83 ug/l

RT: 3.03 min Scan# 400

Delta R.T. -0.05 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 59 Resp: 4804

Ion Ratio Lower Upper



59 100


57 0.0 8.2 12.2#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0





#13

Acrolein

Concen: 6.19 ug/l

RT: 2.28 min Scan# 278

Delta R.T. -0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 56 Resp: 778

Ion Ratio Lower Upper

56 100

55 62.9 57.0 85.4



#15

Acrylonitrile

Concen: 0.50 ug/l

RT: 3.15 min Scan# 420

Delta R.T. -0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09




Tgt Ion: 53 Resp: 744


Ion Ratio Lower Upper

53 100

52 88.3 66.7 100.1

51 28.2 29.9 44.9#





#16

Acetone

Concen: 42.48 ug/l

RT: 2.45 min Scan# 305

Delta R.T. -0.00 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 43 Resp: 62096

Ion Ratio Lower Upper

43 100

58 27.5 22.1 33.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.45

40000

30000

20000



10000

0

Time-->

2.40

2.50



#17

Carbon Disulfide

Concen: 0.60 ug/l

RT: 2.57 min Scan# 325

Delta R.T. -0.00 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 76 Resp: 4047

Ion Ratio Lower Upper

76 100

78 6.6 6.9 10.3#



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.57

2500

2000

1500

1000

500

0


Time-->


2.50

2.55

2.60

2.65





#18

Methyl Acetate

Concen: 0.38 ug/l

RT: 2.78 min Scan# 360

Delta R.T. 0.00 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 43 Resp: 1572

Ion Ratio Lower Upper




43 100

74 26.7 16.7 25.1#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0



#20

Methylene Chloride

Concen: 0.71 ug/l

RT: 2.86 min Scan# 372

Delta R.T. -0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Tgt Ion: 84 Resp: 2016


Ion Ratio Lower Upper

84 100

49 134.3 107.8 161.6

51 33.1 32.8 49.2

86 52.4 52.5 78.7#






Abundance Ion 83.90 (83.60 to 84.60): VX0

Ion 48.90 (48.60 to 49.60): VX0

Ion 50.90 (50.60 to 51.60): VX0

Ion 85.90 (85.60 to 86.60): VX0





#21

trans-1,2-Dichloroethene

Concen: 0.21 ug/l

RT: 3.17 min Scan# 424

Delta R.T. 0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampled :


Tot Ion: 96 Resp: 574

Ion Ratio Lower Upper

96 100

61 92.2 117.0 175.4#




98 47.5 52.4 78.6#



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#25

2-Butanone

Concen: 6.52 ug/l

RT: 4.71 min Scan# 676

Delta R.T. 0.00 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 43 Resp: 14150

Ion Ratio Lower Upper



43 100


72 25.2 18.5 27.7



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0





#29

Tetrahydrofuran

Concen: 0.51 ug/l

RT: 5.17 min Scan# 752

Delta R.T. 0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampled :


Tot Ion: 42 Resp: 715

Ion Ratio Lower Upper

42 100

72 51.5 32.0 48.0#




71 53.6 30.1 45.1#



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.59 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 65 Resp: 178549

Ion Ratio Lower Upper



65 100


67 51.4 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.87 min Scan# 1030
Delta R.T. 0.00 min
Lab File: VX002899.D
Acq: 26 Jun 2018 16:09

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion:114 Resp: 365086
Ion Ratio Lower Upper
114 100
63 19.9 0.0 41.4
88 14.6 0.0 30.0




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VX
Ion 87.90 (87.60 to 88.60): VX






#35
Dibromofluoromethane
Concen: 48.81 ug/l
RT: 5.51 min Scan# 807
Delta R.T. 0.00 min
Lab File: VX002899.D
Acq: 26 Jun 2018 16:09

Tgt Ion:113 Resp: 141114
Ion Ratio Lower Upper
113 100
111 101.3 81.4 122.0
192 24.0 19.1 28.7



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.39 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.13 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 75 Resp: 17564

Ion Ratio Lower Upper

75 100

110 9.7 18.1 54.4#

77 0.5 24.3 36.5#





Abundance Ion 74.90 (74.60 to 75.60): VX002859.D (-845) (-)

Ion 109.90 (109.60 to 110.60): VX002859.D (-845) (-)

Ion 76.90 (76.60 to 77.60): VX002859.D (-845) (-)

Time-->



#38

Carbon Tetrachloride

Concen: 5.25 ug/l

RT: 5.67 min Scan# 834

Delta R.T. -0.12 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09


Tgt Ion: 117 Resp: 23082

Ion Ratio Lower Upper

117 100

119 5.2 77.4 116.0#

121 0.0 24.4 36.6#





Abundance Ion 116.80 (116.50 to 117.50): VX002859.D (-841) (-)

Ion 118.80 (118.50 to 119.50): VX002859.D (-841) (-)

Ion 120.80 (120.50 to 121.50): VX002859.D (-841) (-)

Time-->





#49

1,4-Dioxane

Concen: 0.77 ug/l

RT: 7.59 min Scan# 1149

Delta R.T. -0.16 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampled :


Tot Ion: 88 Resp: 56

Ion Ratio Lower Upper

88 100

43 0.0 29.8 44.6#




58 51.8 57.1 85.7#



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 48.68 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 98 Resp: 513935

Ion Ratio Lower Upper



98 100


100 63.6 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#51

4-Methyl-2-Pentanone

Concen: 0.30 ug/l

RT: 8.65 min Scan# 1323

Delta R.T. 0.00 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X





ClientSampled :

Tot Ion: 43 Resp: 1365

Ion Ratio Lower Upper

43 100

58 35.9 28.6 42.8



#59

2-Hexanone

Concen: 0.21 ug/l

RT: 9.50 min Scan# 1462

Delta R.T. -0.00 min

Lab File: VX002899.D




Acq: 26 Jun 2018 16:09


Tgt Ion: 43 Resp: 723

Ion Ratio Lower Upper

43 100

58 48.8 24.6 73.6









#62
4-Bromofluorobenzene
Concen: 44.46 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX002899.D
Acq: 26 Jun 2018 16:09

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion: 95 Resp: 181017
Ion Ratio Lower Upper
95 100
174 95.5 0.0 187.4
176 93.1 0.0 181.0




Abundance Ion 94.90 (94.60 to 95.60): VX0
Ion 173.80 (173.50 to 174.50): V
Ion 175.80 (175.50 to 176.50): V






#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX002899.D
Acq: 26 Jun 2018 16:09

Tgt Ion: 117 Resp: 337863
Ion Ratio Lower Upper
117 100
82 54.1 45.0 67.4
119 32.8 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.08 min Scan# 1885

Delta R.T. -0.01 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampled :


Tot Ion:152 Resp: 190940

Ion Ratio Lower Upper

152 100

115 54.8 40.1 120.2




150 156.1 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 27.69 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.16 min

Lab File: VX002899.D


Acq: 26 Jun 2018 16:09

Tgt Ion: 75 Resp: 95765

Ion Ratio Lower Upper



75 100


77 1.3 22.8 68.3#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#81

trans-1,4-Dichloro-2-butene

Concen: 78.30 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.06 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09

Instrument :

MSVOA_X

ClientSampled :


Tot Ion: 75 Resp: 95765

Ion Ratio Lower Upper

75 100

53 0.2 86.8 130.2#




89 0.0 38.2 57.2#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#95

Naphthalene

Concen: 0.25 ug/l

RT: 13.84 min Scan# 2174

Delta R.T. -0.00 min

Lab File: VX002899.D

Acq: 26 Jun 2018 16:09


Tgt Ion: 128 Resp: 3218

Ion Ratio Lower Upper

128 100

127 16.9 10.4 15.6#



129 13.5 8.6 13.0#




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 0.21 ug/l
 RT: 14.02 min Scan# 2204
 Delta R.T. 0.00 min
 Lab File: VX002899.D
 Acq: 26 Jun 2018 16:09

Instrument :
 MSVOA_X
 ClientSampleId :

Tot Ion:180 Resp: 1033
 Ion Ratio Lower Upper
 180 100
 182 80.6 48.0 144.0
 145 50.6 14.8 44.3#

