

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062618\
 Data File : VX002914.D
 Acq On : 27 Jun 2018 00:36
 Operator : JC/MD
 Sample : J3707-08 50X
 Misc : 5.97g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021SB480-13-15-180622

Quant Time: Jun 27 04:34:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251517	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	352129	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327973	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	183933	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	169984	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	5.51	113	134979	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
50) Toluene-d8	8.72	98	497040	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.14	95	174897	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
Target Compounds						
10) Methyl Iodide	2.52	142	1589	6.72	ug/l	# 93
27) cis-1,2-Dichloroethene	4.60	96	9764	3.22	ug/l	78
44) Trichloroethene	7.21	130	1832	0.55	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062618\
Data File : VX002914.D
Acq On : 27 Jun 2018 00:36
Operator : JC/MD
Sample : J3707-08 50X
Misc : 5.97µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021SB480-13-15-180622

Quant Time: Jun 27 04:34:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

