


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062618\
 Data File : VX002890.D
 Acq On : 26 Jun 2018 09:36
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Title : SW846 8260
 Last Update : Mon Jun 25 14:38:05 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	3612	PASS
75	95	30	60	51.6	9080	PASS
95	95	100	100	100.0	17590	PASS
96	95	5	9	7.0	1238	PASS
173	174	0.00	2	1.7	273	PASS
174	95	50	100	93.1	16384	PASS
175	174	5	9	7.8	1274	PASS
176	174	95	101	98.9	16197	PASS
177	176	5	9	6.6	1067	PASS