

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062619\
 Data File : VX010451.D
 Acq On : 26 Jun 2019 10:42
 Operator : JC/SP
 Sample : K3539-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-6-25-19

Quant Time: Jun 27 02:15:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	275039	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	421703	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	381571	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	175126	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	129927	50.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.76%	
35) Dibromofluoromethane	5.50	113	128012	49.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.22%	
50) Toluene-d8	8.71	98	496925	50.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.13	95	166168	46.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062619\
Data File : VX010451.D
Acq On : 26 Jun 2019 10:42
Operator : JC/SP
Sample : K3539-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-6-25-19

Quant Time: Jun 27 02:15:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

