

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062619\
 Data File : VX010465.D
 Acq On : 26 Jun 2019 16:09
 Operator : JC/SP
 Sample : K3527-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-CPMW09S-M24W1-062219

Quant Time: Jun 27 03:11:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387237	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	355497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167762	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	118469	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
35) Dibromofluoromethane	5.50	113	118627	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
50) Toluene-d8	8.71	98	462161	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	11.13	95	156282	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
Target Compounds						
16) Acetone	2.44	43	4561	3.853	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	1344	0.495	ug/l	90
30) Chloroform	5.21	83	3797	0.942	ug/l	99
64) Tetrachloroethene	9.34	164	43264	15.046	ug/l	99
65) Chlorobenzene	10.13	112	1915	0.265	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062619\
Data File : VX010465.D
Acq On : 26 Jun 2019 16:09
Operator : JC/SP
Sample : K3527-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-CPMW09S-M24W1-062219

Quant Time: Jun 27 03:11:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

