

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062619\
 Data File : VX010449.D
 Acq On : 26 Jun 2019 09:55
 Operator : JC/SP
 Sample : VX0626WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0626WBL01

Quant Time: Jun 27 01:24:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	244612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	376632	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	334807	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144963	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114636	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
35) Dibromofluoromethane	5.49	113	114985	49.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.80%	
50) Toluene-d8	8.71	98	439691	49.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.13	95	142803	44.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.74%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062619\
Data File : VX010449.D
Acq On : 26 Jun 2019 09:55
Operator : JC/SP
Sample : VX0626WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0626WBL01

Quant Time: Jun 27 01:24:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010449.D

Acq: 26 Jun 2019 09:55

Instrument :

MSVOA_X

ClientSampled :


VX0626WBL01

Tot Ion:168 Resp: 244612

Ion Ratio Lower Upper

168 100

99 46.6 35.5 53.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 49.983 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX010449.D


Acq: 26 Jun 2019 09:55

Tgt Ion: 65 Resp: 114636

Ion Ratio Lower Upper

65 100

67 55.0 0.0 110.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

50000

40000

30000


20000


10000

0

Time-->

5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX010449.D

Acq: 26 Jun 2019 09:55

Instrument :

MSVOA_X

ClientSampleId :

VX0626WBL01


Tot Ion:114 Resp: 376632

Ion Ratio Lower Upper

114 100

63 17.2 0.0 33.8




88 14.0 0.0 27.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.899 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX010449.D

Acq: 26 Jun 2019 09:55


Tgt Ion:113 Resp: 114985

Ion Ratio Lower Upper

113 100

111 103.0 81.2 121.8



192 23.4 18.6 27.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.874 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010449.D

Acq: 26 Jun 2019 09:55

Instrument :

MSVOA_X

ClientSampleId :


VX0626WBL01

Tot Ion: 98 Resp: 439691

Ion Ratio Lower Upper

98 100

100 64.0 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.71

300000



200000

100000

0

8.60 8.70 8.80

Time-->



#62

4-Bromofluorobenzene

Concen: 44.866 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX010449.D

Acq: 26 Jun 2019 09:55


Tgt Ion: 95 Resp: 142803

Ion Ratio Lower Upper

95 100

174 95.2 0.0 187.6

176 92.4 0.0 184.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

150000


100000


50000

0

11.10 11.20

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX010449.D
Acq: 26 Jun 2019 09:55

Instrument :
MSVOA_X
ClientSampleId :
VX0626WBL01

Tot Ion:117 Resp: 334807
Ion Ratio Lower Upper
117 100
82 50.3 41.2 61.8
119 32.7 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX010449.D
Acq: 26 Jun 2019 09:55

Tgt Ion:152 Resp: 144963
Ion Ratio Lower Upper
152 100
115 54.0 38.6 115.7
150 159.3 0.0 347.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

