

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062620\
 Data File : VX016983.D
 Acq On : 26 Jun 2020 15:07
 Operator : JC/SP
 Sample : L3097-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB02-20200623

Quant Time: Jun 29 06:14:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 26 08:34:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	215278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	388643	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	391104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	196698	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	176433	54.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.54%	
35) Dibromofluoromethane	5.47	113	130152	50.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.40%	
50) Toluene-d8	8.70	98	488669	51.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.88%	
62) 4-Bromofluorobenzene	11.12	95	201631	55.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.84%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062620\
Data File : VX016983.D
Acq On : 26 Jun 2020 15:07
Operator : JC/SP
Sample : L3097-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB02-20200623

Quant Time: Jun 29 06:14:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 26 08:34:16 2020
Response via : Initial Calibration

