

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062620\
 Data File : VX016985.D
 Acq On : 26 Jun 2020 15:52
 Operator : JC/SP
 Sample : L3117-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-200624

Quant Time: Jun 27 04:48:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 26 08:34:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	221674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	406464	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	408317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	182715	54.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.16%	
35) Dibromofluoromethane	5.47	113	135358	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	8.70	98	508370	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	11.12	95	217834	57.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.50%	
Target Compounds						
16) Acetone	2.42	43	4907	3.750	ug/l	94
21) trans-1,2-Dichloroethene	3.14	96	1838	0.662	ug/l	88
27) cis-1,2-Dichloroethene	4.56	96	28258	9.211	ug/l	86
44) Trichloroethene	7.19	130	839	0.272	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062620\
Data File : VX016985.D
Acq On : 26 Jun 2020 15:52
Operator : JC/SP
Sample : L3117-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-200624

Quant Time: Jun 27 04:48:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 26 08:34:16 2020
Response via : Initial Calibration

