

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062620\
 Data File : VX016992.D
 Acq On : 26 Jun 2020 18:31
 Operator : JC/SP
 Sample : L3112-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB01-20200623

Quant Time: Jun 29 06:40:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 26 08:34:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	246567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	449351	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	456166	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	238423	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	186972	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
35) Dibromofluoromethane	5.47	113	148087	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
50) Toluene-d8	8.70	98	554629	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
62) 4-Bromofluorobenzene	11.12	95	236911	56.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.64%	
Target Compounds						
4) Vinyl Chloride	1.39	62	4816	1.557	ug/l	97
16) Acetone	2.42	43	9105	6.256	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	40509	11.872	ug/l	88
44) Trichloroethene	7.19	130	202906	59.488	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062620\
Data File : VX016992.D
Acq On : 26 Jun 2020 18:31
Operator : JC/SP
Sample : L3112-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB01-20200623

Quant Time: Jun 29 06:40:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 26 08:34:16 2020
Response via : Initial Calibration

