


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062620\
 Data File : VX016960.D
 Acq On : 25 Jun 2020 18:07
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062620W.M
 Title : SW846 8260
 Last Update : Fri Jun 26 08:34:16 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	11269	PASS
75	95	30	60	51.5	29182	PASS
95	95	100	100	100.0	56618	PASS
96	95	5	9	6.8	3855	PASS
173	174	0.00	2	1.3	554	PASS
174	95	50	100	78.2	44285	PASS
175	174	5	9	7.5	3339	PASS
176	174	95	101	99.4	44018	PASS
177	176	5	9	6.7	2935	PASS