

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
 Data File : VX036332.D
 Acq On : 26 Jun 2023 16:29
 Operator : JC/MD
 Sample : 03319-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 27 04:25:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	240948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	120924	58.595	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 117.180%#		
35) Dibromofluoromethane	5.385	113	87816	55.274	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.540%		
50) Toluene-d8	8.647	98	326815	55.075	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 110.160%		
62) 4-Bromofluorobenzene	11.079	95	107058	51.221	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 102.440%		
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	1947	2.207	ug/l #	64
20) Methylene Chloride	2.788	84	872	0.459	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
Data File : VX036332.D
Acq On : 26 Jun 2023 16:29
Operator : JC/MD
Sample : 03319-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 27 04:25:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

