

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
 Data File : VX036333.D
 Acq On : 26 Jun 2023 16:53
 Operator : JC/MD
 Sample : 03319-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 27 04:25:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	195222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377442	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	352474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174081	62.208	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 124.420%#	
35) Dibromofluoromethane	5.385	113	128680	51.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.400%	
50) Toluene-d8	8.647	98	460406	49.530	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.060%	
62) 4-Bromofluorobenzene	11.079	95	170892	52.194	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.380%	
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	2167	1.812	ug/l #	88
20) Methylene Chloride	2.794	84	1309	0.508	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
Data File : VX036333.D
Acq On : 26 Jun 2023 16:53
Operator : JC/MD
Sample : 03319-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 27 04:25:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

