

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
 Data File : VX036334.D
 Acq On : 26 Jun 2023 17:16
 Operator : JC/MD
 Sample : 03319-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jun 27 04:25:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	183256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	358504	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	335665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164195	62.506	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 125.020%#	
35) Dibromofluoromethane	5.385	113	123506	52.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.500%	
50) Toluene-d8	8.647	98	436874	49.481	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.960%	
62) 4-Bromofluorobenzene	11.079	95	161165	51.823	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.640%	
Target Compounds						
16) Acetone	2.404	43	2741	2.441	ug/l #	81
20) Methylene Chloride	2.794	84	1523	0.630	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
Data File : VX036334.D
Acq On : 26 Jun 2023 17:16
Operator : JC/MD
Sample : 03319-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jun 27 04:25:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

