

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
 Data File : VX036336.D
 Acq On : 26 Jun 2023 18:03
 Operator : JC/MD
 Sample : 03339-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jun 27 04:26:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	175493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	350414	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	323972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164978	65.583	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 131.160%#	
35) Dibromofluoromethane	5.385	113	121927	52.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.540%	
50) Toluene-d8	8.647	98	422027	48.903	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.800%	
62) 4-Bromofluorobenzene	11.079	95	158336	52.089	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.180%	
Target Compounds						Qvalue
16) Acetone	2.392	43	3235	3.009	ug/l	97
78) n-propylbenzene	11.305	91	4991	0.378	ug/l	95
95) Naphthalene	13.774	128	6723	0.657	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062623\
Data File : VX036336.D
Acq On : 26 Jun 2023 18:03
Operator : JC/MD
Sample : 03339-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jun 27 04:26:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

