

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062624\
 Data File : VX042047.D
 Acq On : 26 Jun 2024 12:49
 Operator : JC/MD
 Sample : VX0626WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0626WBL01

Quant Time: Jun 27 03:42:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	159968	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	244958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79141	42.903	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.800%
35) Dibromofluoromethane	5.385	113	76590	44.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.140%
50) Toluene-d8	8.647	98	280705	45.971	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.940%#
62) 4-Bromofluorobenzene	11.079	95	102921	45.726	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062624\
Data File : VX042047.D
Acq On : 26 Jun 2024 12:49
Operator : JC/MD
Sample : VX0626WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0626WBL01

Quant Time: Jun 27 03:42:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

