

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062624\
 Data File : VX042055.D
 Acq On : 26 Jun 2024 16:00
 Operator : JC/MD
 Sample : P3069-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A1285

Quant Time: Jun 27 03:44:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	151448	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	229237	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69451	39.768	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.540%
35) Dibromofluoromethane	5.385	113	69563	42.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.540%
50) Toluene-d8	8.647	98	261184	45.708	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.420%#
62) 4-Bromofluorobenzene	11.079	95	101507	48.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.380%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5279	6.884	ug/l	93
20) Methylene Chloride	2.782	84	1872	1.115	ug/l	# 88
52) Toluene	8.720	92	10329	2.784	ug/l	99
67) Ethyl Benzene	10.195	91	30799	4.235	ug/l	97
68) m/p-Xylenes	10.299	106	18831	6.414	ug/l	89
69) o-Xylene	10.640	106	10264	3.531	ug/l	87
70) Styrene	10.659	104	18508	3.797	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	8950	1.559	ug/l	89
84) 1,2,4-Trimethylbenzene	11.750	105	29548	5.131	ug/l	87
95) Naphthalene	13.774	128	63209	8.411	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062624\
Data File : VX042055.D
Acq On : 26 Jun 2024 16:00
Operator : JC/MD
Sample : P3069-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A1285

Quant Time: Jun 27 03:44:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

