

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062624\
 Data File : VX042057.D
 Acq On : 26 Jun 2024 16:47
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 27 03:45:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	142244	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	217066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67023	40.861	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.720%
35) Dibromofluoromethane	5.391	113	66211	42.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.980%
50) Toluene-d8	8.647	98	250647	46.323	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.640%
62) 4-Bromofluorobenzene	11.079	95	97968	49.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.240%
Target Compounds						
95) Naphthalene	13.780	128	13189	2.382	ug/l	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062624\
Data File : VX042057.D
Acq On : 26 Jun 2024 16:47
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 27 03:45:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

