

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002934.D
 Acq On : 27 Jun 2018 14:18
 Operator : JC/MD
 Sample : J3707-19
 Misc : 5.79g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021SB489-21-23-180622

Quant Time: Jul 06 17:14:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	239104	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	338192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	324263	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195710	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	162445	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
35) Dibromofluoromethane	5.51	113	125440	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
50) Toluene-d8	8.72	98	483735	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.14	95	179808	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	131693	45.73	ug/l	83
65) Chlorobenzene	10.14	112	6413	0.78	ug/l	95
91) 1,2-Dichlorobenzene	12.40	146	3780	0.55	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002934.D
Acq On : 27 Jun 2018 14:18
Operator : JC/MD
Sample : J3707-19
Misc : 5.79µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021SB489-21-23-180622

Quant Time: Jul 06 17:14:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

