

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002935.D
 Acq On : 27 Jun 2018 14:44
 Operator : JC/MD
 Sample : J3707-01
 Misc : 5.49g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP021SB488-21-23-180622

Manual Integrations
 APPROVED

MMDadoda
 6/28/2018 4:47:53 PM

Quant Time: Jul 06 17:17:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	248662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	348247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	335645	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	169362	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
35) Dibromofluoromethane	5.51	113	128898	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
50) Toluene-d8	8.72	98	498068	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.14	95	191450	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
Target Compounds						
4) Vinyl Chloride	1.41	62	75317	26.78	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	2120	0.71	ug/l	80
88) 1,4-Dichlorobenzene	12.10	146	3841m	0.51	ug/l	
89) n-Butylbenzene	12.39	91	2746	0.25	ug/l #	72
91) 1,2-Dichlorobenzene	12.40	146	10090	1.37	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002935.D
Acq On : 27 Jun 2018 14:44
Operator : JC/MD
Sample : J3707-01
Misc : 5.49µ/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP021SB488-21-23-180622

Manual Integrations
APPROVED

MMDadoda
6/28/2018 4:47:53 PM

Quant Time: Jul 06 17:17:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

