

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002964.D
 Acq On : 28 Jun 2018 04:25
 Operator : JC/MD
 Sample : J3718-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-24

Quant Time: Jun 28 15:34:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	372127	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	205595	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	179383	48.99	ug/l	0.00
Spiked Amount			Recovery	=	97.98%	
35) Dibromofluoromethane	5.51	113	145393	49.34	ug/l	0.00
Spiked Amount			Recovery	=	98.68%	
50) Toluene-d8	8.72	98	523771	48.67	ug/l	0.00
Spiked Amount			Recovery	=	97.34%	
62) 4-Bromofluorobenzene	11.14	95	193261	46.57	ug/l	0.00
Spiked Amount			Recovery	=	93.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	7.46	83	1775	0.37	ug/l	# 93
40) Benzene	6.15	78	9625	0.81	ug/l	96
65) Chlorobenzene	10.14	112	4657	0.53	ug/l	99
67) Ethyl Benzene	10.26	91	3933	0.27	ug/l	94
73) Isopropylbenzene	11.02	105	40900	2.96	ug/l	98
78) n-propylbenzene	11.37	91	5980	0.38	ug/l	98
83) tert-Butylbenzene	11.77	119	10617	0.92	ug/l	94
85) sec-Butylbenzene	11.95	105	38281	2.75	ug/l	82
88) 1,4-Dichlorobenzene	12.10	146	3171	0.43	ug/l	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002964.D
Acq On : 28 Jun 2018 04:25
Operator : JC/MD
Sample : J3718-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-24

Quant Time: Jun 28 15:34:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

