

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002966.D
 Acq On : 28 Jun 2018 05:18
 Operator : JC/MD
 Sample : J3718-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-34

Quant Time: Jun 28 15:39:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	385024	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	361741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	211657	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	187130	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
35) Dibromofluoromethane	5.51	113	149912	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
50) Toluene-d8	8.72	98	540055	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
62) 4-Bromofluorobenzene	11.14	95	199263	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	2.19	74	2139	1.13	ug/l	81
11) Tert butyl alcohol	3.07	59	17816	23.61	ug/l	100
16) Acetone	2.45	43	9498	6.05	ug/l	93
19) Methyl tert-butyl Ether	3.21	73	173179	17.82	ug/l	99
22) Diisopropyl ether	3.87	45	14293	1.40	ug/l #	91
39) Methylcyclohexane	7.47	83	3070	0.62	ug/l #	88
65) Chlorobenzene	10.14	112	7376	0.80	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002966.D
Acq On : 28 Jun 2018 05:18
Operator : JC/MD
Sample : J3718-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-34

Quant Time: Jun 28 15:39:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

