

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002967.D
 Acq On : 28 Jun 2018 05:45
 Operator : JC/MD
 Sample : J3718-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-29

Quant Time: Jun 28 07:44:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272184	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	385423	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	361203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	209195	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	183539	48.79	ug/l	0.00
Spiked Amount			Recovery	=	97.58%	
35) Dibromofluoromethane	5.51	113	147469	48.32	ug/l	0.00
Spiked Amount			Recovery	=	96.64%	
50) Toluene-d8	8.72	98	543365	48.75	ug/l	0.00
Spiked Amount			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	11.14	95	194337	45.21	ug/l	0.00
Spiked Amount			Recovery	=	90.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.51	142	3497	7.23	ug/l	# 96
11) Tert butyl alcohol	3.07	59	16903	22.55	ug/l	95
16) Acetone	2.45	43	15416	9.89	ug/l	95
19) Methyl tert-butyl Ether	3.21	73	106400	11.03	ug/l	100
22) Diisopropyl ether	3.88	45	12449	1.23	ug/l	92
65) Chlorobenzene	10.14	112	159628	17.40	ug/l	98
88) 1,4-Dichlorobenzene	12.10	146	22043	2.93	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002967.D
Acq On : 28 Jun 2018 05:45
Operator : JC/MD
Sample : J3718-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-29

Quant Time: Jun 28 07:44:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

